

# Dr's Cam

dars

0

M

# njoy Mvireless & Auto H-ocus

Witeless Autofocus Intraoral Camera with Mouse

GoođÓrs

Do not Touch ((r+1))

#### Intraoral Camera

C-WHT6 -IC-DCAM70 -IC-DCAU71 -IC-WHC50 -IC-F7AU81

Curing Light

-CL-DL10

-CL-AT24

-CL-DLT30 -CL-DC20

by Wireless & Auto Focus!

Wireless Autofocus Intraoral Camera with Mouse

### 1 Features of Dr's cam



......

Π

#### High speed digital wireless

High speed digital wireless provide more clear live image without interruption.

#### 2) Wireless receiver •

Compact wireless receiver without any power supply.

#### Macro to Full arch

Camera is able to focus from macro mode (close) to full arch.


#### Auto focus AF

Can get clear image with auto focus function in oral cavity.

HD image HD

Can get high definition image without distortion.

Shadowless Lighting Rounded 12pcs LEDs make shadowless lighting.



280° twist head Easy to get approach anywhere in oral cavity.

#### OLED screen

Bright and clear OLED screen indicates various device status.

#### Adjustable LED brightness

To avoid image quality loss by light reflection, can adjust LED brightness.

#### 3 Color Impressions

Can change live image color impression (Original, Warm, Cool)

#### Software compatibility

Can be use camera with most of dental imaging software. It would need twain driver for some of dental imaging software.

## 2 Specification

 $\odot$ 

Model name		IC-DCAM70(Wireless model)
Camera	Operation voltage	3.7V
	Size	Approx. 255 x 30 x 28 mm
	Weight	Approx. 89g
	Resolution	HD(720P)
	Sensor/Lens type	CMOS / Liquid lens
Charger	Operation voltage	5V
	Size	Approx. 110 x 105 x 68 mm
	147 1 1 1	
	Weight	Approx. 100g
Disposable cover	Size	Approx. 100g Approx. 32 x 100 mm
Disposable cover		

Model name		IC-DCAU71(Wired model)
Camera	Operation voltage	5V
	Size	Approx. 202 x 30 x 28 mm
	Weight	Approx. 56g
	Resolution	HD(720P)
	Sensor/Lens type	CMOS / Liquid lens
USB cable	Length	Approx. 3M
Disposable cover	Size	Approx. 32 x 100 mm
	Quantity	Approx. 100 pcs by weight x 2ea
	Raw material	Polyethylene

## 3 Minimum System Requirement

PC system:	2.0 Ghz CPU, 1GB RAM
OS:	Windows XP service pack 3
USB port:	USB 2.0





Upper images are take by real product



Main View

Live View

Print View



4 DrsViewer2

Image Viewer included.

#### Good Doctors Co., Ltd.

#208, B-dong, 283 Bupyeong-daero, Bupyeong-gu (Woolim Lions Valley, Cheongchoen-dong), Incheon, 21315, Republic of Korea Tel: +82 32 424 6325 Fax: +82 32 424 6326 E-mail : good@gooddrs.com Website : www.gooddrs.com



Wireless

Wired